

MOISTUREMETER D

LOCAL MEASUREMENTS
OF TISSUE WATER



MOISTUREMETER D

- Measures tissue water content at different layers of skin locally and non-invasively
- Measurement depth determined by the probe size
- Data collection wirelessly into Delfin Modular Core software



MOISTUREMETERD COMPACT

- Fully portable
- Excellent in hospitals, public and private health care centers
- Skin water content measurement by integrated probe
- Built-in pressure sensor for user independent measurements
- Displays tissue water content in percent (0 - 100 %)
- Rechargeable battery
- Wireless data collection with Delfin Modular Core software



MOISTUREMETERD

Extremely useful in detecting physiological reactions in the skin and fat layer caused by tissue fluid changes

- Increase or reduction of edema
- Increase of blood flow
- Irritation reaction
- Water changes caused by enhanced active
- Ingredients penetration
- Effect of treatments (medication, lymph therapy, personal care products)
- Long term or transient effect



MOISTUREMETERD DEVICES

MEASUREMENT PRINCIPLE

Interaction of 300 MHz low power electromagnetic waves with tissue water molecules




Portion of the EM energy is absorbed by tissue water

Reflected wave contains information on tissue water content

From reflected wave information **tissue dielectric constant (TDC)** can be determined.



Tissue dielectric constant is proportional to tissue water content



Probe L50

Probe S15



MOISTUREMETERD

WATER SPECIFIC MEASUREMENT USING 300 MHZ RADIOFREQUENCY

1. Extracellular space contains free water and bound (or motionally restricted) water molecules on the surface of macromolecules.
2. Dielectric constant of free water depends on radiofrequency.
Electrical properties of free and bound water present in tissues change with frequency. At around 300 MHz, electrical properties of free and bound water are quite practically equal i.e. whole tissue water compartment is measured, (Pennock, Schwan: J Phys Chem, 1969)
3. MoistureMeterD devices (at 300 MHz) measure the tissue total water content (consisting of free and bound water).



MOISTUREMETERD

SIMULATION STUDY FOR SKIN WATER AND SOLID COMPONENT USING WATER-ETHANOL MIXTURES

- Dielectric constants of pure ethanol and pure water are 22 and 78.5, respectively
- Dielectric constants of mixtures of water and ethanol vary between 22 and 78.5
- Five different mixtures (percentage water volume content, V%) were prepared:


V% = 0 % (pure ethanol)

V% = 25 %

V% = 50 %

V% = 80 %

V% = 100 % (pure water)



- The dielectric constants of mixtures were measured with the MoistureMeterD




MOISTUREMETERD

DIELECTRIC CONSTANT AND WATER CONTENT

- Dielectric constant of water = 78.5
- Dielectric constant of air = 1
- Dielectric constant of skin, typically 20 – 40
- Dielectric constant of human fat ≈ 10
- Tissue water dominates the measurement value due to its exceptionally high dielectric constant
- The deeper the measurement depth, the more adipose tissue included in the measurement field of the probes

⇒ the dielectric constant value decreases



MOISTUREMETERD

SELECTING THE PROBE SIZE

Probe	Effective measuring depth	Measures down to	Application examples
XS5	0.5 mm	2.5 mm	Cosmetic products, transdermal drug penetration
S15	1.5 mm	3 mm	Irritation reactions, wound healing, burns, transdermal drug penetration
M25	2.5 mm	5 mm	Blood circulation, lymphedema, postoperative edema
L50	5 mm	8 mm	Subcutaneous tissue metabolism, cellulite treatment

XS5



S15



M25



L50



OBS! MoistureMeterD Compact has an integrated probe corresponding to the MoistureMeterD probe M25!




MOISTUREMETERD

DETERMINING THE MEASUREMENT DEPTH

An experimental two-layer model



1. A cylindrical chamber with a micrometer-controlled movable acrylic piston filled with water
2. When the piston is in contact with the probe, the MoistureMeterD measures only the piston (dielectric constant = 3)
3. The piston is gradually moved away from the probe and the MoistureMeterD starts to measure the total dielectric constant of the combination of water and the piston and the value increases
4. When the piston is far enough from the probe, the MoistureMeterD measures only water (dielectric constant = 78.5)



MOISTURE METER D

DETERMINING THE MEASUREMENT DEPTH



Dielectric constant vs. thickness of the water layer



MOISTURE METER D

DETERMINING THE MEASUREMENT DEPTH

Dielectric constant vs. thickness of the water layer



MOISTUREMETERD

APPLICATIONS


- Secondary lymphedema
 - related to cancer, surgery, drugs, trauma
- Cutaneous edema, tissue fluid status
 - Postoperative fluid status, sepsis follow-up
 - Effect of haemodialysis
 - Ischaemic disorders of extremities
 - Swollen tissue problems
- Cancer medical drug therapy
 - Radiation reactions
 - Drug response in skin
- Burns, thermal injury, wound healing
 - Assessment of burn depth
 - Periwound healing
 - Inflammation-induced edema
- Dermatology, Irritant and allergic skin reactions, drug reactions in skin
- R&D of pharmaceutical products
 - Effect of topical or systemic drugs and ingredients on skin
- R&D of cosmetic and personal care industry
 - Effects of cellulite treatment
 - In-vitro measurement of water containing samples



MOISTUREMETERD

EXAMPLE

Measurement of irritant- exposed skin by a dielectric technique, probe S15




Measurement of oedema in irritant-exposed skin by a dielectric technique
Miettinen M., Mönkkönen J., Lahtinen M.R., Nuutinen J. and Lahtinen T.,
Skin Res Technol, Vol 12: 235-240, 2006.

MOISTUREMETERD

EXAMPLE

Unilateral postmastectomy arm lymphedema



Segment volume technique



Dielectric technique with the MoistureMeterD

Mayrovitz, ICBS 2005:

The dielectric method may serve as

- (1) a rapid quantitative assessment procedure for documenting lymphedema, and
- (2) early detection of incipient lymphedema that is not yet clinically observable



MOISTUREMETERD

EXAMPLE

Changes in abdominal subcutaneous fat fluid content with rapid weight loss

9 week's very-low-calorie diet and weight maintenance period 6, 9 and 12 months

	Baseline	9 weeks	6 months	9 months	12 months
Weight (kg)	102.5 ± 12.8	86.9 ± 10.4	86.2 ± 11.3	87.7 ± 10.8	88.2 ± 12.4
SAF dielectric constant	23.3 ± 2.3	25.0 ± 2.1	24.3 ± 2.4	27.5 ± 2.0	27.8 ± 1.9

SAF=subcutaneous abdominal fat

Conclusion: Dielectric constant of SAF continued increasing while the weight was maintained. This may describe the better metabolism and increased blood volume in the subcutaneous fat.





MOISTUREMETER D

EXAMPLE

Interstitial fluid removed during haemodialysis treatment, N=7

Dielectric constant was measured with the MoistureMeter D, probe M25




Conclusion: Change of dielectric constant correlates well with the fluid removal from the patient. The change in the dielectric measurement was 4-fold compared with the circumferential measurement.



MOISTUREMETERD

EXAMPLE

Dielectric constant of skin and subcutaneous fat to assess fluid changes after cardiac surgery, measurement probe S15



Conclusion: The measured tissue dielectric constant indicates edema caused by fluid gain after surgery. The dielectric constant is easier and faster to measure than weigh the bed confined patient.



DELFIN CONTACT DETAILS

HEAD OFFICE

Delfin Technologies Ltd.
P.O. Box 1199
70211 Kuopio
FINLAND

Phone +358 50 911 1199

info@delfintech.com



WWW.DELFINTECH.COM

